




Mastercool[®]
"World Class Quality"

COMBUSTIBLE GAS LEAK DETECTOR

**TROUBLESHOOTING
(55750)**



Battery Check: Always check the batteries before any kind of repair. Weak batteries can create false readings.

	ISSUE	SOLUTION
1.	No keypad reaction	Replace the keypad (55750-KEY)
2.	No leak readings	<ul style="list-style-type: none">• The sensor may need to be replaced (55750-SEN) (Sensors usually last around 20 hours)• The probe may need to be replaced (55750-PR)
3.	When the unit is not responding to any function	<ul style="list-style-type: none">• Look for loose or broken wires on the PC board• May need to change the PC board (55750-PC)
4.	LED lights do not show any readings	Replace the LED light (55100-LED)
5.	Broken wire on PC board connecting to battery	Change the battery wire (55750-WIRE)

REPLACEMENT PARTS

DESCRIPTION	PART #
SENSOR	55750-SEN
PROBE	55750-PR
PC BOARD	55750-PC
KEY PAD	55750-KEY
LED LIGHT	55100-LED
BATTERY WIRE	55750-WIRE